

# CAURPOS *Candida auris* Control 5E3 cp/mL

## Instructions for Use

Product Overview	
Name	<i>Candida auris</i> Control 5E3 cp/mL
Part Number	CAURPOS
Matrix	Transport Media
Analyte	<i>Candida auris</i>
Fill Volume	1.0mL
Number of vials	15
Storage Temperature	-20°C or below
Regulatory Status	For In Vitro Diagnostic Use

### Product Description

The *Candida auris* Control 5E3 cp/mL contains whole, intact heat-inactivated *Candida auris*. This control is designed to evaluate the full molecular diagnostic workflow, including nucleic acid extraction, amplification, and detection. Each lot is formulated in a proprietary matrix containing preservatives and tested using digital PCR quantification, ensuring lot-to-lot consistency. The product is non-infectious, stable under recommended storage conditions, and suitable for use as a daily run control in a laboratory setting.

### Intended Use

The *Candida auris* Control 5E3 cp/mL is an external quality control intended for use with molecular diagnostic assays for the detection of *Candida auris* DNA. This control is designed to evaluate assay performance, including test precision, lot-to-lot consistency, limit of detection, and operator technique. Routine use of the control allows laboratories to monitor the accuracy and reliability of their molecular workflow.

### Principles of the Procedure

The *Candida auris* Control 5E3 cp/mL contains whole, heat-inactivated *Candida auris* organism formulated in a transport media with preservatives. The organism is structurally intact, allowing the control to be processed in parallel with patient specimens to assess the entire molecular diagnostic workflow, including nucleic acid extraction, amplification, and detection. This control must be extracted and tested using the same procedures applied to clinical samples. When processed alongside clinical samples, it serves as a full-process control for verifying assay performance, monitoring reagent integrity, and assessing operator technique.

End users are responsible for establishing expected performance ranges for their specific assay system. The *Candida auris* Control 5E3 cp/mL is not intended to replace assay-specific internal or external controls provided by test kit manufacturers.

The control can also be used in analytical validation studies, such as limit of detection, linearity, or precision assessments. It is suitable for dilution using the FWDX Transport Media Dilution Matrix (TMNEG).

### Storage and Handling

The *Candida auris* Control 5E3 cp/mL should be stored at -20°C or below. Thaw *Candida auris* Control 5E3 cp/mL at room temperature and vortex and centrifuge briefly prior to use. Once thawed, *Candida auris* Control 5E3 cp/mL is stable for 24 hours when stored at 2-8°C. Product can be frozen and reused a maximum of four (4) times. After the fourth freeze/thaw, any remaining materials must be disposed of.

Do not use *Candida auris* Control 5E3 cp/mL beyond the expiration date.

### Limitations

For in vitro diagnostic use. Intended for professional use only. Each laboratory must establish its own standard operating procedures and instructions.

# CAURPOS Candida auris Control 5E3 cp/mL Instructions for Use

## Warnings and Precautions

The Candida auris Control 5E3 cp/mL contains heat-inactivated whole organism. Although non-infectious, the product should be handled as potentially biohazardous. Always use universal precautions when handling the control and any associated materials. Do not pipette by mouth. Avoid smoking, eating, or drinking in areas where controls or clinical specimens are handled.

Wear appropriate personal protective equipment (PPE), including gloves, lab coats, and eye protection when handling this product. Avoid inhalation, ingestion, or direct contact with skin and eyes. In case of skin or eye contact, flush thoroughly with water.

In case of inadequate ventilation, use respiratory protection. Wash contaminated clothing before reuse. Clean spills immediately using a freshly prepared 0.5% sodium hypochlorite (bleach) solution or equivalent disinfectant.

Dispose of all waste materials, including used vials and consumables, as if they contain infectious agents, in accordance with local, regional, and national biohazard regulations. Do not use product if packaging is damaged, leaking, or past the expiration date.

## Symbols



Catalog Number



Lot Number



Expiration Date



Upper Limit of Temperature



Caution



For In Vitro Diagnostic Use



Positive Control



Instructions For Use



Manufacturer